



BADAN NARKOTIKA NASIONAL
PUSAT LABORATORIUM NARKOTIKA

FORMULIR CEK KONDISI ALAT HPLC

Nama Instrumen :

Model/Tipe :

Angka Suhu/Kelembaban*) :

	Hari-1 (tanggal) :	Hari-2 (tanggal) :	Hari-3 (tanggal) :
Suhu			
Kelembaban			

KONSENTRASI	ULANGAN (RIPITASI)	WAKTU RETENSI (Rt)	AREA
	I		
	II		
	III		
	IV		
	V		
MEAN			
SD			
%RSD			
KEBERTERIMAAN		< 2%	
SIMPULAN			
PARAF ANALIS			

*) Catat angka Suhu/Kelembaban diluar batas keberterimaan selama 3 hari berturut-turut